



Radiant™ Green Lo-ROX qPCR Kit

Catalog No.	Pack Size and Concentration	Components (Volume)
QS1001	100 x 20µl Reactions, 2X	1 x 1mL
QS1005	500 x 20µl Reactions, 2X	5 x 1mL
QS1020	2000 x 20µl Reactions, 2X	20 x 1mL
QS1050	5000 x 20µl Reactions, 2X	50 x 1mL
QS1100	10,000 x 20µl Reactions, 2X	100 x 1mL

Description

Alkali Scientific offers Radiant™ qPCR Kits, a high-performance portfolio of 2X ready-to-use real-time PCR reagents designed for best-in-class quantitative PCR. Radiant™ qPCR Kits are engineered for robust real-time PCR with earlier quantification cycle values (Ct), industry-leading sensitivity (increased limit of detection) and exceptional speed (rapid extension rates). The proprietary buffer system allows for highly efficient amplification of GC-rich and AT-rich sequences in addition to dramatic improvements in PCR sensitivity in low-copy assays and PCR conditions conducive to ultra-fast amplification. The Radiant™ Green Lo-ROX qPCR Kit utilizes a proprietary, Green non-inhibitory intercalating dye for robust, fluorescent detection.

- Next-generation PCR buffer formulations for maximum PCR efficiency and reaction speed.
- Novel hot-start chemistry for improved specificity and sensitivity.
- Broad range detection for increased reliability in low-copy assays
- Versatile 2X mixes, suitable for both ultra-fast and standard cycling methods

Instrument Compatibility

The Radiant™ Green Lo-ROX qPCR Kit has been optimized for use optimized for use with intercalating-dye based real-time PCR on qPCR instruments that require a standard, low concentration of ROX such as Applied Biosystems® 7500, 7500 FAST, ViiA7 and Stratagene (Agilent)® Mx4000, Mx3000P, Mx3005P. In addition, the Radiant™ Green Lo-ROX qPCR Kit is optimal for use on qPCR instruments that do not require the use of ROX for data normalization such as Bio-Rad Opticon™, Opticon2™, MiniOpticon, Chromo4™, CFX96, CFX384, Roche LightCycler®480, Qiagen Rotor-Gene™ 3000 & 6000, Eppendorf Mastercycler®ep realplex, Cepheid SmartCycler™, and Illumina® Eco™.

Components

Radiant™ Green 2X qPCR Mix Lo-ROX is comprised of novel hot-start *Taq* DNA polymerase, proprietary real-time PCR buffer and Green intercalating dye, optimized MgCl₂, dNTPs, ROX reference dye, enhancers, and stabilizers.

Storage

Radiant™ Green Lo-ROX qPCR Kit is shipped on blue or dry ice and should be stored at –20°C upon receipt. Excessive freeze/thawing should be avoided. When stored as specified, Radiant™ Green Lo-ROX qPCR Kit is stable for 12 months from date of receipt. The Kit may also be stored at 4°C for 1 month.

Important Considerations

- Use primer-design software, such as Primer3 (<http://frodo.wi.mit.edu/primer3/>) or visual OMP™ (<http://dnasoftware.com/>). Primers should have a melting temperature (T_m) of approximately 60°C.
- Optimal amplicon length should be 80bp-200bp, and should not exceed 400bp.

Reaction setup

1. Before starting, briefly vortex Radiant™ Green 2X qPCR Mix Lo-ROX.
2. Prepare a PCR master mix based on following table:

Component	Volume / Reaction	Final Concentration
Radiant™ Green 2X qPCR Mix Lo-ROX	10 µl	1X
Forward Primer (10µM)	0.8 µl	400 nM
Reverse Primer (10µM)	0.8 µl	400 nM
Template DNA	<100ng cDNA, <1µg genomic	Variable
Nuclease-free water	Variable	
Total Reaction Volume*	20 µl	

* For alternative total reaction volumes (eg. 25 µl), scale all components proportionally and maintain final concentrations.

3. Program the instrument using the following conditions, acquiring data on the appropriate channel:

Cycles	Temperature & Time	Notes
1	95°C, 2 min. activation for cDNA	3 minutes for genomic DNA
40	a. 95°C, 5 sec (Denaturation) b. 60°C to 65°C, 20 to 30 seconds (Annealing/Extension)	Do not exceed 30 seconds and do not use temps below 60°C

Quality Control

Radiant™ Green Lo-ROX qPCR Kit is tested extensively for robust activity, processivity, efficiency, heat activation, sensitivity, absence of nuclease contamination and absence of nucleic acid contamination. Radiant™ Green Lo-ROX qPCR Kit is manufactured under a comprehensive quality management system, following ISO 9001:2008 standards.

Limitations of Use

This product is intended for research purposes only and is not intended for any animal or human therapeutic use.

Technical Support

For Trouble-shooting and Technical Guidance, please contact us at tech@alkalisci.com and provide PCR reaction conditions, cycling parameters, amplicon size, and screen grabs (amplification traces) if possible.

Alkali Scientific guarantees the performance of all products in the manner described in our product literature. The purchaser must determine the suitability of the product for its particular use. Alkali Scientific shall not be liable for any direct, indirect, consequential or incidental damages arising out of the use, the result of use or the liability to use this product. For additional information, contact: Alkali Scientific Inc. 5370 NW 35th Terrace, Suite 112. Fort Lauderdale, FL 33309 USA; Telephone number: (1) 954-586-4170; Fax: (1) 954-590-8322; Email: sales@alkalisci.com